

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÈQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod lisem 129/2, 171 02 Praha 8 - Troja

CERTIFICATE

No.: 1220226

Product: Baby breathing monitor

Type: BM-02

Rating: 2 x battery AA-LR6 (2 x 1,5 V DC)

Ordering firm: JABLOTRON ALARMS a.s.
Pod Skalkou 4567/33, 466 01 Jablonec nad Nisou, Czech Republic

Manufacturer: JABLOTRON ALARMS a.s.
Pod Skalkou 4567/33, 466 01 Jablonec nad Nisou, Czech Republic

Factories: JabloPCB s.r.o.
U Přehrady 5129/67, 466 04 Jablonec nad Nisou, Czech Republic

Trade mark:

The test results are stated in the test-report No.: 220505-01/01 of: 26.04.2022, 220505-01/02 of: 04.04.2022

A sample of the product was found to be in conformity with:

ČSN EN 60601-1 ed. 2:2007+A1:2014+A2:2022, ČSN EN 60601-1-2 ed. 3:2016+A1:2021,
ČSN EN 60601-1-6 ed. 3:2010+A1:2015+A2:2021, ČSN EN 60601-1-11 ed. 2:2016+A1:2021

Other data:

The validity of the certificate is limited to: 08.05.2025

09.05.2022

Prague

Mgr. Miroslav Sedláček
Head of Certification Body



220505-01